

**MIZORAM STATE HEALTH SOCIETY  
STATE PROGRAMME MANAGEMENT UNIT  
OFFICE OF THE MISSION DIRECTOR  
NATIONAL HEALTH MISSION  
MIZORAM : AIZAWL**

No.D.21013/1/2018-NHM/MSHS/SPMU/PRO

Dated Aizawl the 9<sup>th</sup> February 2018

**SEALED SHORT TENDER NOTICE**

Sealed Short Tenders are hereby invited as per the schedule annexed to this notice on behalf of the Chairman, Mizoram State Health Society from reputed and bonafide Manufacturers or their only authorized dealers, for **Rate Contract of Medical Equipments, Instruments etc under NHM, Mizoram** which will be received by the Mission Director, NHM, Health & FW Department, Dinthar, Aizawl, Mizoram on or before **27-02-2018 (Tuesday) upto 12:00 Noon** . Tenders are to be opened by the Technical Committee under the chairmanship of Mission Director, NHM, Mizoram, Aizawl or his authorized representatives at **1:00 PM** on the same day. The Tenderers or their representatives may also be present at the time of opening of the Quotations, if they so desire.

Terms and Conditions-

- 1) Tenders shall be submitted in duplicate and accordingly marked "**Original**" and "**Duplicate**" as the case may be ,on the cover of the envelope. The outer cover of the tender shall be superscribed as "**Rate Contract of Medical Equipments, Instruments etc under NHM, Mizoram**" addressed to the Mission Director, National Health Mission, Dinthar, Aizawl clearly indicating full postal address with fax no.of the tender/sender.
- 2) Tenders shall be submitted in the prescribed form only which can be obtained from the office of Mission Director, National Health Mission, Dinthar, Aizawl, Mizoram on working days.
- 3) Tenders shall be accompanied with the Earnest Money Deposit of 2% of quoted value in the form of Demand Draft/Deposit at Call in original, from any Nationalized Bank of India or Postal Saving Certificate duly pledged in favor of Mission Director, National Health Mission, Mizoram, Aizawl.
- 4) Manufacturers submitting their tenders directly shall attach photo copies of their Manufacturing Licence, CIN (Corporate Identity Number) and other relevant permissions/licences required for such manufacturing duly attested by a Gazetted Officer or Notary. Original Certificates shall be produced at the time of opening of the tenders/quotations, if demanded.
- 5) Tenders received from firms and individuals who are not authorized dealers but only authorized to quote on behalf of the manufacturer or dealer shall be accepted. Authorized dealers shall attach Authorization or Distributorship or Dealership, as the case may be from the concerned Manufacturers. A Gazetted Officer or Notary,

- however, shall duly attest such photo copies. Original Certificate shall be produced at the time of opening of the tenders/quotations,if demanded
- 6) Quoted rates in terms of Indian Rupees shall be considered inclusive of all Taxes.If there are more than one brand, the separate rates for such brands with specific details should be mentioned.
  - 7) Warranty period should be mentioned clearly and availability of After-Sale- Services facilities during warranty period.
  - 8) F.O.R. destination is Central Medical Store, Zemabawk, Aizawl shall be clearly mentioned.
  - 9) Soft copy of tender document shall be enclosed alongwith other documents.
  - 10) Tender must be accompanied by the following documents-
    - a) Photo copy of GST Registration Certificate duly attested by a Gazetted Officer. Original Certificates should be produced at the time of opening of the tenders/quotations, if demanded.
    - b) Photo copy of Scheduled Tribe Certificate duly attesteds by a Gazetted Officer or Notary (for Tribal tenderer). Original Certificates should be produced at the time of opening of the tenders/quotations,if demanded.
    - c) Non-Judicial Adhesive Stamp worth at least Rs.7.50 should be affixed (for non-tribal tenderer)
  - 11) The tenderer shall submit Attested or Notorized Photo copy of valid and upto date ISO/GMP Certificate
  - 12) All pages of the tender papers should be printed,signed,stamped and dated; hand written quotations shall not be entertained. Any page or pages not signed, stamped and dated will be marked violated clause at the time of opening of tender.
  - 13) Tenders must enclose detail specifications of the items. They must also enclose relevant documents including catalogue /literature of the quoted items.
  - 14) A successful tenderer has to execute an appropriate Deed of Agreement within 15 (fifteen)days from the date of communication of acceptance of his/her quoted rate (s) by fax, failing which the tender will become invalid.
  - 15) Once the rate is accepted and communicated to the successful firm, the same will remain valid during the supply period of the agreement with the approved supplier and no price escalation will be permissible during the valid period, which is normally for a period of 12 momths and is extendable, if mutually agreed upon by both the parties i.e. State Health Mission and concerned approved supplier, subject to the recommendation of the Purchase Committee, Mizoram State Health Society, as the case may be.
  - 16) A successful tenderer whose tender has been accepted shall furnish Security Deposit of 5% (five percent) of the value of the materials required to be supplied, at the time of execution of the agreement,in the form of Demand Draft/Fixed Deposited in any Nationalized Bank of India or Postal Savings Certificate, as the case may be, duly pledged in favor of Mission Director, NHM, Government of Mizoram, Aizawl.
  - 17) A successful tenderer shall supply the items within the stipulated time. If the supply cannot be completed within the stipulated period without any valid and justifiable reasons with prior intimation, the Society shall be at liberty to procure the materials from the next lowest tenderer and the differences in prices, if any will be realized from the Security Deposit of the successful and approved tenderer without assigning any reason thereof.

- 18) The Society shall not be bound to accept the lowest rate or any of the tenders even if valid, and reserves the right to accept or reject any tender without assigning any reason thereof/based on previous performance. The Society may give preference to an item of reputed brand/company with ISO specification.
- 19) In case of any legal dispute/litigations arising out of this Tender Notice or its terms, the local courts shall have the jurisdiction as per the relevant laws in force.
- 20) All payments will be made only after the materials are received in full and in good condition and after completion of installation and satisfactory commissioning.
- 21) If the above items are violated or unfulfilled, tender shall be rejected at the time of opening of the tender.



**(Dr.LALROZAMA)IAS**  
Mission Director , NHM  
Mizoram: Aizawl

Memo No.D.21013/1/2018-NHM/MSHS/SPMU/PRO

Dated Aizawl the 9<sup>TH</sup> February 2018

Copy to:

1. P.S to Hon'ble Minister, Health & Family Welfare Department, Govt. of Mizoram
2. P.S to Secretary, Health & Family Welfare Department, Govt. of Mizoram
3. Principal Director, Health & Family Welfare Department for information.
4. Director of Health Services, Mizoram, Aizawl for information
5. Director of Hospital and Medical Education, Aizawl for information.
6. The Editor\_\_\_\_\_with a request for publication of the above caption in your newspaper for atleast 2 consecutive days.
7. Guard File



Mission Director, NHM  
Mizoram: Aizawl

## LIST OF MEDICAL EQUIPMENTS , INSTRUMENTS etc UNDER NHM

Annexure

SI No	Particular	Specifications	Rate in Figure	Rate in Word
1	Removable Stretcher on Trolley	Stretcher on Trolley Removable / Detachable stainless steel top Should have strong tubular framework of trolley mounted on four 15cms swivelling castors		
2	Crush Trolley	Stainless steel material with 4 no. of castor Should have minimum 4 shelves		
3	WheelChair	Had bearing capacity – 100 kg Seat width between 16 – 20 inches Should have fixed armrest and footrest Rear wheel brake support Rexine / Leather like polyester fiber material for the back & bottom.		
4	Refridgerator (for storing medicines & Chemos)	Minimum 192 ltrs		
5	Finger Pulse Oximeter	Should have plethismographic wave form with numeric display for SPO2 and Heart rate on LCD/TFT display <ul style="list-style-type: none"> <li>- Should have a SPO2 range of 0 to 100%</li> <li>- Should have SPO2 accuracy of <math>\pm 2\%</math></li> <li>- Should provide bar graph for pulse strength</li> <li>- Audio and visual alarm for both upper and lower SPO2, Heart rate</li> <li>- Should provide with adult reusable finger probe with technology from standard reputed companies</li> <li>- Beep sound and alarm sound should have separate volume control</li> <li>- Should have a minimum of 2 hours back-up time</li> <li>- Should be a portable, light weight and desktop model</li> <li>- Should work with input 200 to 240Vac 50 Hz supply</li> <li>- Should have safety certificate from competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced alongwith the bid</li> </ul>		
6	Oxygen Flowmeter	Should be plugin type, rail mounted with humidifier Can adjust accurate prescribed flow from the gas outlet for the patient Flow meter should be made of copper Available for 0-15 / 0-10 LPM Range.		
7	Suction Machine	Volt 230 vac Rating of motor continuous Suction Bottle Capacity : 2 x 2000ml minimum (with safety valve) Guage : 0 to 760 mm Hg		

SI No	Particular	Specifications	Rate in Figure	Rate in Word
		<p>Pump : Oil lubricates lottery pump</p> <p>Suction Tubings : ID 7 mm, 5m long and non-collapsible</p> <p>Should have air tights</p> <p>Should have noiseless operation</p> <p>Should provide filter to absorb moisture and water particles into the rotor</p> <p>Should have an external provision for topping up lubricant</p> <p>Should be well-designed, cabinet made of mild steel powder coated</p> <p>Should bear ISI mark</p>		
8	Bedside Screen	<p>Size approx – 1680mm(H) x 2450mm(H)</p> <p>Stainless steel frame mounted on</p> <p>Supplied with hook, spring wire &amp; curtains</p>		
9	Kidney Tray	<p>Size – 10 – 12” (inches)</p> <p>- Stainless steel material</p>		
10	Cardiac Table	<p>Rectangular framework mounted on 5 cms castors</p> <p>- adjustable by gear handle</p> <p>- 760 x 400mm (LxB) minimum size</p>		
11	Dressing Drum Stainless Steel	Stainless steel material ,9x6 inch		
12	Dressing Drum Stainless Steel	Stainless steel material (15x12 inch)		
13	Allis Tissue Forcep	Stainless steel material 6” length		
14	Micro Halsted Mosquito Artery Forcep	Stainless steel material 6” length		
15	Cheatle Forcep	Stainless steel material 6-8 “ long		
16	Needle Holder	<p>Stainless steel material</p> <p>Diameter 5 mm</p> <p>Straight Jaw</p> <p>Handle O type</p>		
17	Nebulizer	<p>Should be lightweight, portable, compact, a dust filter, able to deliver a flow rate / 7 lpm, air pressure / 35 psi, check valve to protect the device against contamination due to backward inhalation, compatible for continuous use</p> <p>&amp; works on 200-240Vac/50Hz.</p>		
18	Glocometer	<p>Should be a hand held meter, reading range/linearity, use electrochemical technology, measuring unit in mg/dl, wide operating temperature, life time replacement offer, easy code entry technique, Battery should be replaceable without using any tools, facility to ensure accuracy of measurements.</p>		
19	Glocometer Strips	Should be able to use capillary blood samples, a minimum 4 months shelf life after opening the strip vial.		
20	Sahlis Haemoglobinomet	<p>Square H.B. Tubes : 2, H.B. Pipettes : 1 Amber colour glass bottle : 1 Glass Stirrer : 1, Dropper</p>		

SI No	Particular	Specifications	Rate in Figure	Rate in Word
	er	with teat : 1 Cleaning Brush : 1, Spare latex tubing one meter :1		
21	N/10 Hydrochloric Acid Reagent			
22	Urine Albumin & Sugar Test Kit			
23	Blood Grouping reagent			
24	Ultra Sonography Machine	System must be state of art model with super computer signal processing and clinically proven imaging technology.		
25	Radiant warmer	Reflector patches to cover the sensor are recommended in order to provide an accurate reading.		
26	Hand-held Fetal Doppler	LED/LCD Display for Display of Fetal Heart Rate, Heart Rate Range available, Doppler Probe should be Light Weight with Holder Facility when not in use.		
27	Delivery Tray	Sterile Labor & Delivery Surgical Tray		
28	Stethoscope	Chestpiece : Dual head nonchill rim & diaphragm Diaphragm : Tunable diaphragm		
29	Digital BP Instrument			
30	Adult Digital Thermometer	Measuring range, Resolution Accuracy , Response time , Distance-to-spot ratio, Adjustable emissivity  Spectral range Laser pointer		
31	Baby Forehead Thermometer (non-touch)	Non Touch		
32	Paediatric Resuscitator Bag	Volume 250ml with masks of 0 and 1 size delivery		
33	Disposable Sterile Gloves	6 inch		
34	Disposable Utility Gloves	6 inch		
35	Wall Clock with			

SI No	Particular	Specifications	Rate in Figure	Rate in Word
	second stand			
36	Wall Mounted Thermometer	Thermometers are the least accurate, but are also the least-invasive.		
37	Blower/ Convector	<ul style="list-style-type: none"> <li>• Fan Forced Hot Air Circulation</li> <li>• Good Heat settings</li> </ul> Thermostat Control to Maintain Desired Room Temperature		
38	Hub Cutter	Saves Healthcare Professional from Needle Stick Injuries, Non Electrical, No Burning, No Fumes, No Noise, No Smell		
39	Autoclave (Electrical or gas stove) Horizontal	Stand alone bench top autoclave 20L, automatic, chamber diameter ca 25cm, depth ca 45, fit with 5L water reservoir, manual fill, autonomy at least 10 cycles, power shut off upon condition of cycle with 3 removable shelves.		
40	Baby Weighting Scale	Durable standard, Reliable result, Smooth performance		
41	Mucus Extractor (De lee's and adult)	Safe to use Reasonable Low maintenance Quality approved Reliable		
42	Sim's Speculum	Excellent exposure of handling A+ Architects, Confident to handle the products.		
43	Hb Meter	<b>Sample material, Measurement range, speedy Results, Sample volume, Operating temperature, Quality control, Calibration, Storage for HemoCue etc</b>		
44	Disposal Gowns	Comfortable to wear Cost effective rates		
45	Disposal face Masks	Made from soft cloth Double layer mask offering excellent bacterial filtration efficiency Super smooth soft ultrasonically welded mask is most comfortable during use Disposable, Sterile ready for use		
46	Disposal Caps	Non-Woven Fabric, Biodegradable		
47	Protective Eye Cover	More impact-resistant than CR39 plastic, Available with all surface treatments (coatings), Good percent of UV filtering, Light weight, Material is very clear		
48	Disposable Baby Receiving sheet	Good Quality		
49	Disposable Umbilical Cord Clamp	Dimensionally accurate Perfect finish Reliable performance		
50	Disposable Sheet for covering Labor Table	Smooth touch Unmatched quality Biodegradable+		
51	Whole Blood	Good quality & safe		

SI No	Particular	Specifications	Rate in Figure	Rate in Word
	Finger Prick Test for HIV			
52	Manual Vacuum Aspirator	Double valve		
53	Mannequin Mama Breast	Height	Approx 160cm	
		Weight	approx 17kg	
		Material	Silicone	
54	RPR Test Kit			
55	Insulin Syringe	40.U/ml		
56	Chemilumine scene Semi-Auto Analyzer	Analytical module with at least 12 test position Simple operation and rapid processing Result in not more than 90 minutes Power 50-150 watts, frequency 50-60 Htz Should be compact Multi parametric,automated immunoassay system based on Enzyme Linked Fluorescent Assy(ELFA) principles.		
57	Near Vision Chart	(Drum Type)		
58	Rotating Vision Acuity Drum			
59	LED Vision Chart MVA- 66	MVA- 66		
60	Trail Box with Frame			
61	Chalazion Clamp Stainless Steel	(large& small)		
62	Chalazion Scoop Stainless Steel(large & small)	(large& small)		
63	Applanation Tonometer	FAT- T170		
64	Universal Eye Speculum Stainless Steel	Stainless Steel		
65	Snellens Chart			
66	Canula (No.26)Stainless Steel	Stainless Steel		
67	Binomag			
68	Iris Repositor Stainless Steel	Stainless Steel		



SI No	Particular	Specifications	Rate in Figure	Rate in Word
69	Cheatles Forcep Stainless Steel	Stainless Steel		
70	Plain Forcep Stainless Steel	Stainless Steel		
71	Forcep Stainless Steel	Stainless Steel		
72	Torch Light	3 cell		
73	Torch Light LED Stainless Steel	Stainless Steel		
74	Direct Ophthalmoscope Heine Beta 200	Heine Beta 200		
75	Streak Retinoscope Heine Beta 200 Stainless Steel	Heine Beta 200 Stainless Steel		
76	Schiotz Tonometer Stainless Steel	Stainless Steel		
77	Trail Frame Adult / Children			
78	Autokerato - refractometer / S.Korea with table (Matrix Premio KR-10)	(Matrix Premio KR-10)		
79	Chalazion Set			
80	Cataract Set (Titanium)			
81	Lacrimal Probe Stainless Steel (1 set contain 3 size)	(1 set contain 3 size)		
82	Lid Retractor Stainless Steel (Desmarres) 3 different size	(Desmarres) 3 different size		
83	Punctum Dilator ( 1 set contain 3 sizes) Stainless Steel	( 1 set contain 3 sizes) Stainless Steel		
84	Foreign Body Spud & needle Stainless Steel	Stainless Steel		
85	Hand	(14 litres capacity, Brass tank)		

SI No	Particular	Specifications	Rate in Figure	Rate in Word
	Compression Pump			
86	HC Pump accessories			
	a)Rubber Washer	(14 trs capacity)		
	b)Flat Fan Nozzle			
	c)Trigger cut off with pipe			
	d)Air check valve			
	e)Pump barrel assembly			
	f)Sprayer rod	(Long hose sprayer)		
87	CDC Miniature Light Trap			
88	No.3 stainless steel insect pins			
89	No.4 stainless steel insect pins			
90	Pipettes	1ml (Glass)		
91	Pipettes	5ml (Glass)		
92	Pipettes	20ml (Glass)		
93	Pipettes	50ml (Glass)		
94	Sample container	50 ml(Tarson)		
95	Dropper	3ml (Plastic)		
96	Micropipette	10-100ul (Tarson)		
97	Micropipette	100-1000 ul (Tarson)		
98	D-Cell aspirator			
	Intra Occular Lens	+20.00D		
		+20.50D		
		+21.00D		
		+21.50D		
		+22.00D		
		+22.50D		
		+23.00D		
99.	Mosquito Rearing Cage	equivalent 12"x12"x12" or its		

	Atropine Eyeointment	1%		
	Chloramphenicol + Hydrocortisone Eyeointment	2 mg		
	Gentamycin eyedrop	0. w/v		
	Gatifloxacin Eyedrop	0.5% (5mg/ml)		
	Moxifloxacin eyedrop	0.5%		
	Timolol Melate eye drop	0.5%		
	Cyclopentolate eyedrop	1%		
	Antibiotic Steroid Eyedrop (Moxifloxacin with Loteprednol)	0.5%+0.5%		
	Antibiotic Steroid Eyedrop (Moxifloxacin with Dexamethasone)	0.3+0.1%		

